


EXTECH[®]

USER MANUAL

Big Digit Hygro- Thermometer MODEL 445703



1 Introduction

Congratulations on your purchase of the Extech Big Digit Hygro-Thermometer. This professional meter, with proper care, will provide years of safe and reliable service.

2 Operation

1. Remove the protective film from the display.
2. Select the temperature units using the °C/°F button on the rear of the instrument.
3. The upper display indicates temperature in degrees C or F. The lower display indicates relative humidity in %.
4. To show the minimum and maximum readings:
 - Press the MAX/MIN button. The highest value measured since the Reset button was last pressed appears on the display.
 - Press the MAX/MIN button again to display the lowest value measured since the Reset button was last pressed.
 - Press the MAX/MIN button again to return to normal operation.
5. To reset (clear) the MAX/MIN memory:
 - Press the RESET button for 1 second (while in the MAX/MIN display mode).
6. The unit can be wall-mounted or placed on a flat surface using the foldout footing.
7. Replace the battery when the low battery flag appears on the display.

3 Specifications

Temperature range	14 to 140°F (-10 to 60°C)
Temperature Accuracy	±1.0°C (-10 to 50°C) ±1.8°F (14 to 122°F)
Humidity range	10 to 99% RH
Humidity Accuracy	±5% RH (25 to 85% RH; 32 to 122°F [0 to 50°C])
Power supply	2 x 1.5 V 'AAA' battery
Dimensions / Weight	4.3 x 3.9 x 0.78 in. (109 x 99 x 20 mm) 6 oz. (169 g)
Compliance	CE

4 Battery Replacement

Open the battery compartment by sliding the battery cover on the rear of the instrument downward as indicated by the arrow. Replace the 'AAA' batteries observing correct polarity and secure the cover.



Do not dispose of used batteries or rechargeable batteries in household waste.

5 Customer Support

Customer Support Local Telephone List:

<https://support.flir.com/contact>

Returns (RMA):

<https://customer.flir.com/Home>

6 Warranty

Teledyne FLIR warrants this Extech brand instrument to be free of defects in parts and workmanship for two years from date of shipment. To view the full warranty text, please visit the support site, link below.

<https://www.flir.com/support-center/warranty/>

Website

<http://www.flir.com>

Customer support

<http://support.flir.com>

Copyright

© 2025, FLIR Systems, Inc. All rights reserved worldwide.

Disclaimer

Specifications subject to change without further notice. Models and accessories subject to regional market considerations. License procedures may apply. Products described herein may be subject to US Export Regulations. Please refer to exportquestions@flir.com with any questions.

Publ. No.: NAS100269
Release: AB
Commit: 108458
Head: 108511
Language: en-US
Modified: 2025-10-29
Formatted: 2025-10-30

